

Identification of the product

Product Name	G-DiaNota		
Catalogue Number	DDGG-10-L100	Production Date	2016-12
Lot Number	06006L	Expiry Date	2018-12


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DDGG-10-L100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Norovirus I</i>	FAM	28.0 ± 3.3	28.6
Positive Control (Ct/Cp) : <i>Norovirus II</i>	YD	30.0 ± 3.3	29.7
Positive Control (Ct/Cp) : <i>Rotavirus</i>	TR	28.0 ± 3.3	27.8
Internal Control (Ct/Cp)	Cy5	C _t ≤ 36.0	28.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 21/02/2017
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.