

Identification of the product

Product name	G-DiaNota		
Product reference	DDGG-10-L100	Production date	2019-08
Lot number	09009H	Expiry date	2021-08


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DDGG-10-L100
Real Time PCR equipment (Supplier)	Biorad CFX96™

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Norovirus I</i>	FAM	28.0 ± 3.3	27.6
Positive Control (Ct/Cp) : <i>Norovirus II</i>	HEX	30.0 ± 3.3	31.5
Positive Control (Ct/Cp) : <i>Rotavirus</i>	TR	28.0 ± 3.3	25.7
Internal Control (Ct/Cp)	Cy5	Ct ≤ 36.0	31.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <i>R. PALEDELT QA Manager</i>	Signature: 	Date: <i>23 Apr 2020</i>
--	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.