

Identification of the product

Product Name	G-DiaNota		
Catalogue Number	DDGG-10-L100	Production Date	2016-05
Lot Number	2017A10/I	Expiry Date	2018-05


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DDGG-10-L100
Real Time PCR equipment (Supplier)	LightCycler 480 II (ROCHE)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) :	FAM	28.0 ± 3.3	27.3
Positive Control (Ct/Cp) :	YD	30.0 ± 3.3	29.2
Positive Control (Ct/Cp) :	TR	28.0 ± 3.3	26.8
Internal Control (Ct/Cp) :	CY5	≤ 36.0	30.4
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 13/02/2017
--	--	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.