

Identification of the product

Product Name	G-DiaFrag		
Catalogue Number	DDGG-20-L100	Production Date	2016-10
Lot Number	01006J	Expiry Date	2018-10

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : Dientamoeba fragilis	FAM	30.0 ± 3.3	29.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : <i>A. ZALOTECY QA Manager</i>	Signature: 	Date: <i>23 Jun 2017</i>
---	---	--------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.