

# Certificate of Analysis

Product identification			
Product name	G-DiaFrag		
Product reference	DDGG-20-L100	Manufacturing date	2021-06
Lot number	2021F15/I	Expiry date	2023-06

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LihgtCycler 480II (Roche)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp):	FAM	30.0 ± 3.3	32.0	Passed /Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed /Fail

Approval			
	Name & Function	Signature	Date
QC	Philippe Naud QC Specialist		30-07-2021
QA	Lamari Samira QA officer		05 AUG 2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.