

**Identification of the product**

Product Name	G-DiaPara		
Catalogue Number	DDGG-25-L100	Production Date	2016-10
Lot Number	01006J	Expiry Date	2018-10

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Giardia lamblia</i>	FAM	30 ± 3.3	29.4
Positive Control (Ct/Cp) : <i>Entamoeba histolytica</i>	YD	30 ± 3.3	31.3
Positive Control (Ct/Cp) : <i>Cryptosporidium parvum</i>	TR	30 ± 3.3	31.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

<b>Name (function) :</b> D. KUSINDA QA OFFICER	<b>Signature:</b> 	<b>Date:</b> 09/02/2017
------------------------------------------------------	----------------------------------------------------------------------------------------------------------	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.