

**Identification of the product**

Product name	G-DiaPara		
Product reference	DDGG-25-L100	Production date	2017-06
Lot number	02007F	Expiry date	2019-06


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Giardia lamblia</i>	FAM	30 ± 3.3	29.7
Positive Control (Ct/Cp) : <i>Entamoeba histolytica</i>	YD	30 ± 3.3	29.5
Positive Control (Ct/Cp) : <i>Cryptosporidium parvum</i>	TR	30 ± 3.3	30.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function :	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> <b>David Kusinda</b>                      QA Officer                 </div>		16/08/2017

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.