

Identification of the product

Product name	G-DiaPara		
Product reference	DDGG-25-L100	Production date	2020-01
Lot number	04000A	Expiry date	2022-01

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Giardia lamblia</i>	FAM	30 ± 3.3	29.5
Positive Control (Ct/Cp) : <i>Entamoeba histolytica</i>	YD	30 ± 3.3	30.4
Positive Control (Ct/Cp) : <i>Cryptosporidium parvum</i>	TR	30 ± 3.3	30.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Caroline TOUSSAINT QA Officer	Signature: 	Date: 23 JUL 2020
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.