

Identification of the product

Product Name	G-DiaDiff		
Catalog Number	DDGG-30-L100	Production Date	2016-04
Lot Number	16176D-1	Expiry Date	2018-04

Analysis procedure

Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Multiplex 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Clostridium difficile tcd A</i>	31.0 ± 3.3	FAM : 29.7
Positive Control (Ct/Cp) <i>Clostridium difficile tcd B</i>	30.0 ± 3.3	YD : 29.8
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: J. KUSINJA
QA OFFICER 

Date: 07/09/2016