

**Identification of the product**

Product Name	G-DiaBact		
Catalog Number	DDGG-35-L100	Production Date	2016-06
Lot Number	16226F	Expiry Date	2018-06

**Analysis procedure**


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Ro Master Mix 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Salmonella enterica</i>	30.0 ± 3.3	FAM: 28.5
Positive Control (Ct/Cp) <i>Campylobacter jejuni</i>	30.0 ± 3.3	YD: 29.7
Negative Control (Ct/Cp)	Negative	Negative

**Approval**

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: D. KUSINDA  
QA OFFICER 

Date: 23/08/2016