

Identification of the product

Product name	G-DiaBact		
Product reference	DDGG-35-L100	Production date	2019-05
Lot number	2019F25/II	Expiry date	2020-05

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Salmonella enterica</i>	FAM	30 ± 3.3	29.9
Positive Control (Ct/Cp) : <i>Campylobacter jejuni</i>	YD	30 ± 3.3	30.0
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 21 AUG 2019
--	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.