

Identification of the product

Product name	G-DiaShig		
Product reference	DDGG-40-L100	Production date	2018-01
Lot number	2018B08/1	Expiry date	2019-01

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Shigella</i>	FAM	30.0 ± 3.3	28.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <i>R. ZALIEVEUS</i> <i>QA/QC Manager</i>	Signature: 	Date: <i>26 Feb 2018</i>
--	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.