

**Identification of the product**

Product name	G-DiaShig		
Product reference	DDGG-40-L100	Production date	2019-05
Lot number	2019F25/III	Expiry date	2020-05


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Julianna KERESZTES QA Officer	Signature: 	Date: <b>05 SEP 2019</b>
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

