


Product identification			
Product name	H-DiaCMVQ		
Product reference	DDGH-10-L200	Production date	2018-12
Batch number	02008L	Expiry date	2019-12

Analysis procedure	
Volume PCR samples (µL)	Positive control: 5.0 Standards: 5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results			
Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	21.2–24.2	23.0
IQC (c/PCR)	FAM	222–1802	463
EIC DNA (Ct)	CY5	< 36.0	26.7
Standard 1 (Ct/Cp)	FAM	17.9–20.9	20.5
Standard 2 (Ct/Cp)	FAM	21.2–24.2	23.5
Standard 3 (Ct/Cp)	FAM	24.5–27.5	26.7
Standard 4 (Ct/Cp)	FAM	27.8–30.8	30.1
Standard 5 (Ct/Cp)	FAM	31.1–34.1	33.8
Negative Control (Ct/Cp)	N/A	Negative	Negative
Slope	N/A	-3.6 to -3.0	-3.0
Efficiency	N/A	1.9 to 2.1	2.0

Quality Control: CHARLIERE Date: 01/04/2019

Quality Assurance:  Date: 01/04/2019