

Identification of the product

Product Name	H-DiaHSV		
Catalog Number	DDGH-40-L100	Production Date	2016-10
Lot Number	2016J10/I	Expiry Date	2018-03

Analysis procedure


Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference (Supplier)	DMML-D2-Q100 (Diagenode)
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) HSV 1	27.4 ± 3.3	FAM: 28.8
Positive Control (Ct/Cp) HSV 2	26.6 ± 3.3	YD: 27.6
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: D. KUSINDA
 QA OFFICER 

Date: 18/11/2016