

Identification of the product

Product name	H-DiaHSV		
Product reference	DDGH-40-L100	Production date	2017-02
Lot number	2017B17//II	Expiry date	2018-08


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Herpes Simplex Virus 1	FAM	27.4 ± 3.3	26.5
Positive Control (Ct/Cp) : Herpes Simplex Virus 2	YD	26.6 ± 3.3	26.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : D. KUSINDA QA OFFICER	Signature: 	Date: 27/04/2017
---	---	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.