

**Identification of the product**

Product name	H-DiaHSV		
Product reference	DDGH-40-L100	Production date	2017-09
Lot number	2017I20/I	Expiry date	2019-03


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Herpes Simplex Virus 1	FAM	27.4 ± 3.3	27.2
Positive Control (Ct/Cp) : Herpes Simplex Virus 2	YD	26.6 ± 3.3	26.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : <div style="border: 1px solid black; padding: 2px; display: inline-block;">                 David Kusinda QA Officer             </div>	Signature: 	Date: 12/10/2017
--	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.