

Identification of the product

Product name	H-DiaHSV		
Product reference	DDGH-40-L100	Production date	2018-06
Lot number	2018G27/1	Expiry date	2020-06


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Herpes Simplex Virus 1	FAM	27.4 ± 3.3	28.4
Positive Control (Ct/Cp) : Herpes Simplex Virus 2	YD	26.6 ± 3.3	26.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> David Kusinda QA Officer </div>		13/08/2018

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.