

**Identification of the product**

Product name	H-DiaHSV		
Product reference	DDGH-40-L100	Production date	2019-06
Lot number	2019G15/1	Expiry date	2021-06


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	27.4 ± 3.3	27.2
Positive Control (Ct/Cp) :	YD	26.6 ± 3.3	25.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Julianna KERESZTES QA Officer	Signature: 	Date: <b>14 AUG 2019</b>
---	--	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

