

**Identification of the product**

Product name	H-DiaHSV		
Product reference	DDGH-40-L100	Production date	2020-08
Lot number	2020H31/I	Expiry date	2022-08



**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp):	FAM	27.4 ± 3.3	28.0
Positive Control (Ct/Cp):	YD	26.6 ± 3.3	25.6
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function: 	Signature: 	Date: <b>17 -09- 2020</b>
---	--	------------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.