

Identification of the product

Product Name	H-DiaVZV		
Catalogue Number	DDGH-50-L100	Production Date	2019-04
Lot Number	04009D	Expiry Date	2021-04


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : Varicella Zoster Virus	FAM	30.0 ± 3.3	30.8
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : Julianna KERESZTES QA Officer	Signature: 	Date: 11 JUL 2019
--	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.