

Identification of the product

Product name	H-DiaVZV		
Product reference	DDGH-50-L100	Production date	2020-05
Lot number	05000E	Expiry date	2022-05


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	31.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 30 JUL 2020
--	---	-------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.