

**Identification of the product**

Product Name	H-DiaVZV		
Catalogue Number	DDGH-50-L100	Production Date	2018-01
Lot Number	2019C20/II	Expiry Date	2020-01

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC Results by Real Time PCR**

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : Varicella Zoster Virus	FAM	30.0 ± 3.3	30.6
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name (function) : <b>C. Couvrer</b> QA SPECIALIST	Signature: 	Date: 04 APR 2019
---	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

