

Identification of the product

Product Name	M-DiaSero		
Catalogue Number	DDGM-20-L040	Production Date	2016-09
Lot Number	020061	Expiry Date	2018-09


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) N. tot :	FAM	30.0 ± 3.3	30.4
Positive Control (Ct/Cp) N. gB :	FAM	30.0 ± 3.3	31.0
Positive Control (Ct/Cp) N. gC :	FAM	30.0 ± 3.3	30.3
Positive Control (Ct/Cp) N. gA :	FAM	30.0 ± 3.3	29.8
Positive Control (Ct/Cp) N. gW135 :	FAM	30.0 ± 3.3	29.1
Positive Control (Ct/Cp) N. gY :	FAM	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	FAM	Negative	Negative

Approval

Name (function) : D. ROSINDA QA OFFICER	Signature: 	Date: 06/12/2016
--	--	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.