

Identification of the product

Product name	M-DiaSero		
Product reference	DDGM-20-L040	Production date	2017-06
Lot number	03007F	Expiry date	2019-06


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) N. tot :	FAM	30.0 ± 3.3	29.2
Positive Control (Ct/Cp) N. gB :	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp) N. gC :	FAM	30.0 ± 3.3	28.5
Positive Control (Ct/Cp) N. gA :	FAM	30.0 ± 3.3	29.8
Positive Control (Ct/Cp) N. gW135 :	FAM	30.0 ± 3.3	29.2
Positive Control (Ct/Cp) N. gY :	FAM	30.0 ± 3.3	29.9
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	<div style="border: 1px solid black; padding: 2px; display: inline-block;"> David Kusinda QA Officer </div>	Signature: 	Date: 05/09/2017
---------------------	---	---	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.