

Identification of the product

Product name	M-DiaSero		
Product reference	DDGM-20-L040	Production date	2019-09
Lot number	050091	Expiry date	2021-09


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) N. tot :	FAM	30.0 ± 3.3	30.1
Positive Control (Ct/Cp) N. gB :	FAM	30.0 ± 3.3	31.2
Positive Control (Ct/Cp) N. gC :	FAM	30.0 ± 3.3	29.3
Positive Control (Ct/Cp) N. gA :	FAM	30.0 ± 3.3	31.0
Positive Control (Ct/Cp) N. gW135 :	FAM	30.0 ± 3.3	30.1
Positive Control (Ct/Cp) N. gY :	FAM	30.0 ± 3.3	30.0
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> Dominique d'Amico QA Officer </div>		21-11-2019

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.