

Product identification			
Product name	M-DiaSero		
Product reference	DDGM-20-L040	Manufacturing date	2021-04
Lot number	07001D	Expiry date	2023-04

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp) N. tot	FAM	30.0 ± 3.3	29.2	Passed / Fail
Positive Control (Ct/Cp) N. gA	FAM	30.0 ± 3.3	29.5	Passed / Fail
Positive Control (Ct/Cp) N. gB	FAM	30.0 ± 3.3	30.8	Passed / Fail
Positive Control (Ct/Cp) N. gC	FAM	30.0 ± 3.3	28.6	Passed / Fail
Positive Control (Ct/Cp) N. gW135	FAM	30.0 ± 3.3	29.4	Passed / Fail
Positive Control (Ct/Cp) N. gY	FAM	30.0 ± 3.3	29.6	Passed / Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	Philippe Nouel qc Specialist		05-07-2021
QA	Lamari Samira QA officer		09 JUL 2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.