

Identification of the product

Product Name	M-DiaMult		
Catalogue Number	DDGM-30-L100	Production Date	2016-09
Lot Number	2017K09/1	Expiry Date	2018-07


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) <i>Neisseria meningitidis</i>	FAM	28.0 ± 3.3	29.0
Positive Control (Ct/Cp) <i>Streptococcus pneuminae</i>	YD	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) :	Signature:	Date:
David Kusinda QA Officer		07/12/2017

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.