

Identification of the product

Product name	M-DiaMult		
Product reference	DDGM-30-L100	Production date	2019-01
Lot number	2018L04/1	Expiry date	2020-12


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) <i>Neisseria meningitidis</i>	FAM	28.0 ± 3.3	28.4
Positive Control (Ct/Cp) <i>Streptococcus pneumoniae</i>	YD	30.0 ± 3.3	30.4
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
Dominique d'Amico QA Officer		05/02/2019

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.