

Certificate of Analysis

Product identification			
Product name	M-DiaMult		
Product reference	DDGM-30-L100	Manufacturing date	2021-01
Lot number	2021B10/IV	Expiry date	2023-01

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-Q100
Real Time PCR equipment (Supplier)	LightCycler 480II(Roche)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp)	FAM	28.0 ± 3.3	28.2	Passed / Fail
Positive Control (Ct/Cp)	YD	30.0 ± 3.3	28.9	Passed / Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	Philippe Nard QC Specialist		24.03.2021
QA	Carine COUVRUER QA Officer		19/04/2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.

