

Identification of the product

Product name	O-DiaEnt		
Product reference	DDGO-30-L100	Production date	2017-06
Lot number	03007F	Expiry date	2019-06



Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.6
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	 David Kusinda QA Officer	Signature:		Date:	11/08/2017
---------------------	--	------------	--	-------	------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.