

Identification of the product

Product name	O-DiaADV		
Product reference	DDGO-10-L100	Production date	2018-05
Lot number	03008E	Expiry date	2020-05


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-D600
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	28.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <div style="border: 1px solid black; padding: 2px; display: inline-block;"> Dominique d'Amico QA Officer </div>	Signature: 	Date: 06/08/2018
---	--	---------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.