

**Identification of the product**

Product name	O-DiaENT		
Product reference	DDGO-30-L100	Production date	2019-08
Lot number	05009H	Expiry date	2021-08

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function :	Signature:	Date:
Dominique d'Amico QA Officer		<b>22 -11- 2019</b>

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.