

Product identification			
Product name	O-DiaENT		
Product reference	DDGO-30-L100	Manufacturing date	2021-06
Lot number	06001F	Expiry date	2023-06

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp):	FAM	30.0 ± 3.3	30.5	Passed / Fail
Negative Control (Ct/Cp):	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	Philippe Nard le Spécialist		20-08-2021
QA	Caroline TOUSSAINT QA Officer		17 AUG 2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.