

Identification of the product

Product Name	R-DiaBorM		
Catalog Number	DDGR-10-L100	Production Date	2016-05
Lot Number	01006E	Expiry Date	2018-05

Analysis procedure


Volume PCR samples (µL)	5
Volume PCR mix (µL)	25
Master Mix reference (Supplier)	OPTIMA Ro Master Mix 2x DNA
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Specifications	Results
Positive Control (Ct/Cp) <i>Bordetella pertussis</i>	30.0 ± 3.3	FAM: 29.1
Positive Control (Ct/Cp) <i>Bordetella parapertussis</i>	30.0 ± 3.3	YD: 30.3
Negative Control (Ct/Cp)	Negative	Negative

Approval

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

Approved by: D. Kusindt

 QA OFFICER

Date: 21/09/2016