

**Identification of the product**

Product name	R-DiaBorM		
Product reference	DDGR-10-L100	Production date	2017-01
Lot number	02007A	Expiry date	2019-01

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Bordetella pertussis</i>	FAM	30.0 ± 3.3	29.1
Positive Control (Ct/Cp) : <i>Bordetella parapertussis</i>	YD	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : A. Zalkely QA/QC Manager	Signature: 	Date: 20 Nov 2017
--	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.