

Identification of the product

Product name	R-DiaBorM		
Product reference	DDGR-10-L100	Production date	2018-01
Lot number	03008A	Expiry date	2020-01




Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Bordetella pertussis</i>	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp) : <i>Bordetella parapertussis</i>	YD	30.0 ± 3.3	29.6
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
		

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.