

Identification of the product

Product name	R-DiaBorM		
Product reference	DDGR-10-L100	Production date	2020-01
Lot number	05000A	Expiry date	2022-01


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D2-R100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : Bordetella pertussis	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp): Bordetella parapertussis	YD	30.0 ± 3.3	30.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
Caroline TOUSSAINT QA Officer		24 MAR 2020

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.