

Identification of the product

Product name	R-DiaADV		
Product reference	DDGR-15-L100	Production date	2019-03
Lot number	06009C	Expiry date	2021-03



Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function :	Signature:	Date:
		14-03-2019

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.