

**Identification of the product**

Product name	R-DiaADV		
Product reference	DDGR-15-L100	Production date	2020-06
Lot number	07000F	Expiry date	2022-06


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480 (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: <b>25 SEP 2020</b>
--	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.