

Product identification			
Product name	R-DiaADV		
Product reference	DDGR-15-L100	Manufacturing date	2021-02
Lot number	08001B	Expiry date	2023-02

Analysis procedure	
Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC Results				
Testing	Channel	Specification	Results	Status
Positive Control (Ct/Cp) :	FAM	30.0±3.3	30.7	Passed / Fail
Negative Control (Ct/Cp)	N/A	Negative	Negative	Passed / Fail

Approval			
	Name & Function	Signature	Date
QC	CORNETA. QC Manager		30 AUG 2021
QA	Samuelo LAMARCA QA officer		13 SEP 2021

This batch has been released by Quality Assurance in compliance with Diagenode's Quality Management System.