

Identification of the product

Product name	R-DiaADV		
Product reference	DDGR-15-L100	Production date	2018-05
Lot number	2018D09/1	Expiry date	2020-04

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-D5-D100
Real Time PCR equipment (Supplier)	LightCycler480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	30.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : <i>R. ZALÉPELS</i> <i>QA / QC Manager</i>	Signature: 	Date: <i>01 Aug 2018</i>
---	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.