

Identification of the product

Product name	R-DiaFlu		
Product reference	DDGR-30-L100	Production date	2017-07
Lot number	05007G	Expiry date	2019-07

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Human Influenza A</i>	FAM	30.0 ± 3.3	28.8
Positive Control (Ct/Cp) : <i>Human Influenza B</i>	YD	30.0 ± 3.3	30.7
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function	Signature:	Date:
<div style="border: 1px solid black; padding: 2px; display: inline-block;"> David Kusinda QA Officer </div>		03/10/2017

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.