

**Identification of the product**

Product name	R-DiaFlu		
Product reference	DDGR-30-L100	Production date	2019-07
Lot number	08009G	Expiry date	2021-07

**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) : <i>Human Influenza A</i>	FAM	30.0 ± 3.3	28.4
Positive Control (Ct/Cp) : <i>Human Influenza B</i>	YD	30.0 ± 3.3	30.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : <i>Lombert J. QA Officer</i>	Signature: 	Date: <i>25 Sep 2019</i>
---	---	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.

