

Identification of the product

Product Name	R-DiaFlu		
Catalogue Number	DDGR-30-L100	Production Date	2017-01
Lot Number	17017A	Expiry Date	2019-01

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Human Influenza A</i>	FAM	30.0 ± 3.3	30.0
Positive Control (Ct/Cp) : <i>Human Influenza B</i>	YD	30.0 ± 3.3	30.1
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

<b>Name (function) :</b> D. KUSINDA QA OFFICER	<b>Signature:</b> 	<b>Date:</b> 20/02/2017
--	---	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.