

Identification of the product

Product Name	R-DiaRes		
Catalogue Number	DDGR-46-L100	Production Date	2016-11
Lot Number	05006K	Expiry Date	2018-11

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

QC Results by Real Time PCR

Testing	Dye	Specifications	Results
Positive Control (Ct/Cp) : <i>Human Metapneumovirus</i>	FAM	30.0 ± 3.3	27.9
Positive Control (Ct/Cp) : <i>Human Respiratory Syncytial Virus</i>	YD	30.0 ± 3.3	28.2
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name (function) : D. KUSINDA QA OFFICER	Signature: 	Date: 31/01/2017
------------------------------------------------------	--------------------------------------------------------------------------------------------------------	----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.