

**Identification of the product**

|                  |              |                 |         |
|------------------|--------------|-----------------|---------|
| Product Name     | R-DiaRes     |                 |         |
| Catalogue Number | DDGR-46-L100 | Production Date | 2017-09 |
| Lot Number       | 06007I       | Expiry Date     | 2019-09 |


**Analysis procedure**

|                                    |                            |
|------------------------------------|----------------------------|
| Volume PCR samples (µL)            | 5.0                        |
| Volume PCR mix (µL)                | 20.0                       |
| Master Mix reference               | DMML-R4-A100               |
| Real Time PCR equipment (Supplier) | LightCycler 480 II (Roche) |

**QC Results by Real Time PCR**

| Testing  | Dye | Specifications | Results  |
|--|-----|----------------|----------|
| Positive Control (Ct/Cp) :<br><i>Human Metapneumovirus</i>             | FAM | 30.0 ± 3.3     | 29.7     |
| Positive Control (Ct/Cp) :<br><i>Humen Respiratory Syncytial Virus</i> | YD  | 30.0 ± 3.3     | 30.0     |
| Negative Control (Ct/Cp)   | N/A | Negative       | Negative |

**Approval**

|                          |                             |                   |  |              |            |
|--------------------------|-----------------------------|-------------------|--|--------------|------------|
| <b>Name (function) :</b> | David Kusinda<br>QA Officer | <b>Signature:</b> |  | <b>Date:</b> | 22/11/2017 |
|--------------------------|-----------------------------|-------------------|--|--------------|------------|

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.