

**Identification of the product**

Product name	R-DiaRes		
Product reference	DDGR-46-L100	Production date	2019-03
Lot number	07009C	Expiry date	2021-03


**Analysis procedure**

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480 II (Roche)

**QC results**

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	29.4
	YD	30.0 ± 3.3	30.5
Negative Control (Ct/Cp)	N/A	Negative	Negative

**Approval**

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 12 JUL 2019
--	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.