

Identification of the product

Product name	R-DiaRes		
Product reference	DDGR-46-L100	Production date	2019-11
Lot number	08009K	Expiry date	2021-11

Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480II (Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp)	FAM	30.0 ± 3.3	28.6
Positive Control (Ct/Cp)	YD	30.0 ± 3.3	30.0
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Caroline TOUSSAINT QA Officer	Signature: 	Date: 17 JAN 2020
---	--	----------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.