

Identification of the product

Product name	R-DiaRes		
Product reference	DDGR-46-L100	Production date	2020-07
Lot number	09000G	Expiry date	2022-07


Analysis procedure

Volume PCR samples (µL)	5.0
Volume PCR mix (µL)	20.0
Master Mix reference	DMML-R4-A100
Real Time PCR equipment (Supplier)	LightCycler 480II(Roche)

QC results

Testing	Channel	Specifications	Results
Positive Control (Ct/Cp) :	FAM	30.0 ± 3.3	28.2
Positive Control (Ct/Cp):	YD	30.0 ± 3.3	30.3
Negative Control (Ct/Cp)	N/A	Negative	Negative

Approval

Name and function : Carine COUVRUER QA Officer	Signature: 	Date: 22 SEP 2020
------------------------------------------------------	-------------------------------------------------------------------------------------------------	-----------------------------

This lot has been reviewed by Quality Assurance in compliance of Diagenode's Quality System.